



## OFFICE OF COUNTY MAYOR GLENN JACOBS

Procurement Division, 1000 N. Central Street, Suite 100, Knoxville, TN 37917

### ADDENDUM III TO BID 3470 DIGITAL MEDICAL X-RAY EQUIPMENT AND SOFTWARE

**ADDENDUM DATE:** October 12, 2023  
**BUYER:** Robert Mackey  
**ORIGINAL OPENING DATE:** October 25, 2023 at 2:00 PM

**Questions and Answers:**

**Question 1: What is the incoming power for the generator (single phase or three phase)?**

**Answer 1: Single Phase 208-240vAc Unit**

**Question 2: The request shows "3 Rayence Medical Specific Image QC Station or Equivalent" There is only 1 Rayence detector on the bid which only requires 1 QC Station.**

**Answer 2: It appears that it should have been 3 Rayence light weight batteries followed by the specifications for the 1 QC Station.**

**Question 3: What type of equipment is to be removed and is it a floor mounted system or an overhead tube crane system?**

**Answer 3: We currently have a floor mounted tube with a Powersource 483 Phase which is the same as the generator.**

**End of Addendum**

*Robert Mackey*

**Robert Mackey, Buyer**



## **OFFICE OF COUNTY MAYOR GLENN JACOBS**

---

Procurement Division, 1000 N. Central Street, Suite 100, Knoxville, TN 37917